













Homechoice Claria Automated Peritoneal Dialysis Toolkit

Homechoice Claria procedure
Transfer from PRO

Equipment - Check expiry dates and fill out white record book and put supplies on your bed.

Machine preparation

1. Turn machine on at the back -this may take a few minutes
2. You may be asked to put in **Patient Activation Code** go to page 6
3. Machine will read **Hello "Your Name"**
4. Press 
5. Machine will read **Connecting to Network** -this may take a few minutes
6. You may be asked to **Confirm Configurations** go to page 8
7. Machine will read **Weight 0.0KG**
8. Press 
9. Enter your weight using the  and 
10. Press 
11. Enter the second number using the  and 
12. Press 
13. Press 
14. Machine will read **Blood Pressure 000/000**
15. Press 


16. Enter the top number of your blood pressure using the



17. Press



18. Enter the bottom number of your blood pressure using the



19. Press



20. Press




21. Machine will read **Press go to start**


22. You can now set up your machine as normal


If you get any side effects, talk to your doctor, pharmacist or nurse. This includes any possible side effects not listed in the package leaflet. You can also report side effects directly via the Yellow Card Scheme at www.mhra.gov.uk/yellowcard. By reporting side effects you can help provide more information on the safety of this medicine

End of Therapy and Disconnect




1. Machine will now read **End of Therapy**
2. Please record the following in your white book


Press  to obtain **I - Drain Volume**

Press  to obtain **Total UF**

Press  to obtain **Avg Dwell Time**

Press  to obtain **Lost Dwell or Added Dwell**










3. Press 
4. Machine will read **close all clamps**
5. Close twist clamp
6. Close all clamps
7. Press 
8. Machine will read **Disconnect Yourself/Close all clamps**
9. **Apply hand rub**
10. Open Minicap packet on a flat surface
11. **Apply hand rub**
12. Disconnect yourself from the machine
13. Apply new Minicap
14. Press 
15. Machine will read **Remove cassette**
16. Open door
17. Remove cassette
18. Close door



19. Press 
20. Machine will read **Please wait.**
21. Machine will read **Connecting to network**
22. Machine will read **Turn me off**
23. Turn the machine off at the back

Entering your Patient Activation Code

Entering the Patient Activation Code is only required during the first time you use a machine (or have a machine swap)

Your Patient Activation Code is: _____

1. Ensure that the modem is attached and turned on.
2. Turn machine on at the back.
3. Wait for machine to read **Enter Activation Code 000-0000-000** - this may take a few minutes
4. Press 
5. The **O** will flash.
6. Press  to enter the first digit.
7. Press  to move on to the next digit.
8. Repeat until all digits are entered.
9. After the 10th digit is entered press 
10. Press 
11. Machine will read **Connecting to Network** - This may take several minutes
12. Machine will read **Confirm Configurations**
13. Press 
14. Machine will read **Are you "Your name"**
15. If your name is correct, press  (If your name is incorrect, press  and call your Renal Unit.
16. Machine will read **Share: Clinic & Baxter**
17. Press 


18. Machine will read **New program received/Review new program**
19. Press  until all therapy settings have been displayed
20. Machine will read **Is program correct?**
21. Press 
22. Machine will display cycles and dwell time
23. Machine will read **Program accepted by cyclor**
24. Go back to page 2



Invalid Activation Code

If the patient activation code was entered incorrectly

1. Press 
2. Re-enter your patient activation code

Identification rejected

If you pressed  when prompted to confirm your name. Your name may have been incorrect because you entered your patient activation code incorrectly.

1. After a brief delay, **ENTER ACTIVATION CODEDE** will appear. Re-enter your code.
2. After re-entering your patient activation code press 
3. If your name is still not correct, press  and contact your renal unit.







Program not valid

Cause:- The system found a problem with your program settings

Call your renal unit to sort out your program on their system.




Confirming Configurations

If your renal unit has changed any of your program settings then this may happen, following the connection to the network.

1. Machine may read **Confirm Configurations**
2. Press 
3. Machine will read **Are you "Your name"**
4. If your name is correct, press  (If your name is incorrect, press  and call your Renal Unit.
5. Machine will read **Share: Clinic & Baxter**
6. Press 
7. Machine will read **New program received/Review new program**
8. Press  until all therapy settings have been displayed
9. Machine will read **is program correct?**
10. Press 
11. Machine will display cycles and dwell time
12. Machine will read **Program accepted by cyclers**
13. Go back to page 2

Network Communication Failure

- a. The modem is not on and the green power light is not illuminated.
- b. The modem has red lights illuminated
- c. The modem is in a location with poor signal.
- d. The Ethernet cable (yellow cable) is not plugged in correctly
- e. There is an outage with the server

1. Press 
2. Remove all the cables and reinsert all the cables and wait for all the green lights to appear.
3. Press 
4. Machine will read **connecting to network**
5. If the problem re occurs
6. Turn the machine off at the back.
7. When it says **Hello....** Press 
8. The machine will read **Press Go to start** Go back to your procedure

NOTE:- If a network connection is not established, you can continue to perform your therapy. Your data will be transferred once a connection is next available.

User Assistance Information

Keep this information available at all times. It includes important information about Baxter's **Homechoice Claria** APD system and the phone numbers to call for assistance. Your dialysis center will help you enter the information in this section.

1.1 Personal and Cyclor Information

Name:

Patient ID Number:

Patient Activation Code (if applicable):

Cyclor Serial Number:

Cyclor Model Number:

1.2 Numbers to Call for Assistance

Baxter Technical Assistance:

Phone Number:

Dialysis Center:

Phone Number:

Primary Contact Name:

When Available: